

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code 81034

Lot No. 025



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2252	MYCOPLASMA PNEUMONIAE IgG	025	2025-04	6x6
B2253	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	025	2025-04	1
C2252	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	025	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	362	≥ 100
Control Serum (UA/mL)	29,8	23,1-47,5
Negative Control (UA/mL)	$<10,0$	<10

CALIBRATOR


0	6	6	0	2	5	6	8	1	4	2	7	4	3	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM


0	6	6	0	2	5	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 16.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.