

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F24 SHRIMP

Code	88000-F24		
Lot No.	034	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9981	88000-F24 SHRIMP	034	2025-05	2x6
B9980	CALIBRATORE 88000-F24 SHRIMP	034	2025-06	1
C9980	CONTROL SERUM 88000-F24 SHRIMP	034	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,1	7,4-16,1

CALIBRATOR

6 3 6 0 3 4 6 5 1 7 5 5 2 6 7 1 0 0

Label checked with barcode reader YES Operator Signature

636034651755267100
Shrimp
CALIBRATOR 0.250 mL IVD
LOT 034
2025-06
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6 3 6 0 3 4 6 2 7 5 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature

636034627500000001
Shrimp
CONTROL+ 0.450 mL IVD
LOT 034
2025-06
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 17/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.