

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ft3 - CHORUS

Code	86518			
Lot No.	010	2025-02	2024-04	

REAGENTS

Label	Description	Lot.	Exp.	NO
D3300	ft3	010	2025-02	6X6
B3300	CALIBRATOR ft3	010	2025-04	1
C3300	CONTROL SERUM ft3	010	2025-04	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	441	≥ 100
Control Serum (pmol/L)	4,1	1,3-8,1

CALIBRATOR

9	6	7	0	1	0	6	3	2	9	3	9	2	8	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

9	6	7	0	1	0	6	2	5	4	.6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	----	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature B



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 16/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.