



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F20 ALMOND

Code	88000-F20		
Lot No.	034	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D1051	88000-F20 ALMOND	034	2025-05	2x6
B1050	CALIBRATORE 88000-F20 ALMOND	034	2025-06	1
C1050	CONTROL SERUM 88000-F20 ALMOND	034	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,1	7,4-16,1

CALIBRATOR

6	7	3	0	3	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

673034651755267100  
Almond  
CALIBRATOR 0.250 mL IVD  
LOT 034  
 2025-06 2°C-8°C  
DIESSSE Monteriggioni Italy CE B1050

CONTROL SERUM

6	7	3	0	3	4	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

673034627500000001  
Almond  
CONTROL+ 0.450 mL IVD  
LOT 034  
 2025-06 2°C-8°C  
DIESSSE Monteriggioni Italy CE C1050

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*  
DATE *16.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.