

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F35 POTATO

Code	88000-F35		
Lot No.	034	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D1071	88000-F35 POTATO	034	2025-05	2x6
B1070	CALIBRATORE 88000-F35 POTATO	034	2025-06	1
C1070	CONTROL SERUM 88000-F35 POTATO	034	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,1	7,4-16,1

CALIBRATOR

6	6	5	0	3	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

665034651755267100
Potato
CALIBRATOR 0,250 mL IVD
LOT 034
2025-06
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	6	5	0	3	4	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

665034627500000001
Potato
CONTROL+ 0,450 mL IVD
LOT 034
2025-06
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 21/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.