

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME Jo-1 - CHORUS

Code	86030		
Lot No.	033	2025-08	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3261	Jo-1	033	2025-08	6X6
B3261	CALIBRATORE Jo-1	033	2025-08	1
C3261	CONTROL SERUM Jo-1	033	2025-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control Serum (AU/mL)	67,9	28,8-100,0

CALIBRATOR

1	9	7	0	3	3	7	1	2	3	3	3	3	2	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *KS*



CONTROL SERUM

1	9	7	0	3	3	7	2	2	4	1	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *KS*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Bene*

DATE: *23/08/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.