



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code	81078		
Lot No.	027	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2471	VARICELLA IgG	027	2025-04	6x6
B2472	CALIBRATORE VARICELLA IgG	027	2025-04	1
C2471	CONTROL SERUM VARICELLA IgG	027	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control Serum (Index)	3,7	3,1-5,7
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

0	1	4	0	2	7	6	0	0	0	0	0	2	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

014027600000230000
Varicella IgG
CALIBRATOR 0,425 mL IVD
LOT 027
2025-04
DIESE Monteriggioni Italy CE

CONTROL SERUM

0	1	4	0	2	7	6	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

014027631570000001
Varicella IgG
CONTROL 0,425 mL IVD
LOT 027
2025-04
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]
DATE 22/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.