



QUALITY CONTROL CERTIFICATE

KIT NAME TG Extended Range

Code	86110		
Lot No.	024	2025-04	2024-05

REAGENTS

Label	Description	Lot	Exp.	NO
D4050	TG Extended Range	024	2025-04	6x6
B4050	CALIBRATOR TG Extended Range	024	2025-05	1
C4050	CONTROL SERUM TG Extended Range	024	2025-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	350	≥ 100
CONTROL SERUM (ng/ml)	2,3	1,4-3,4

CALIBRATOR

4	1	8	0	2	4	6	6	1	6	3	0	3	0	6	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *Green*



CONTROL SERUM

4	1	8	0	2	4	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Green*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Green*

DATE 24.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.