

DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

### QUALITY CONTROL CERTIFICATE

KIT NAME DIESSSE TPHA-DAT

Code	26035		
LOT.NO	012	2025-02	2024-05

### REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
TPHA/101/V	E4060	TPHA Positive Control	006	2026-01	1
TPHA/102/V	B4060	TPHA Test Cells	008-2	2025-03	2
TPHA/103/V	C4060	TPHA Control Cells	007-2	2025-02	1
TPHA/104/V	D4060	TPHA Diluent	006-2	2025-03	1
TPHA/105/V	F4060	TPHA Negative Control	006	2025-11	1

### CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
Negative Control	Negative	Negative
Positive Control	Positive	Positive

*Evaluation of the method according to the instruction for use.  
This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 24/05/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIALI

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.