

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOSCANA VIRUS IgG-CHORUS

Code 81067

Lot No. 024



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2411	TOSCANA VIRUS IgG	024	2025-04	6X6
B2412	CALIBRATORE TOSCANA VIRUS IgG	024	2025-04	1
C2411	CONTROL SERUM TOSCANA VIRUS IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	518	≥ 100
Control Serum (INDEX)	1,3	0,9-1,6
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR


0	1	3	0	2	4	6	0	0	0	0	0	8	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM

0	1	3	0	2	4	6	0	9	1	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *26.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.