



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F17 HAZELNUT

Code	88000-F17		
Lot No.	036	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9841	88000-F17 HAZELNUT	036	2025-05	2x6
B9840	CALIBRATORE 88000-F17 HAZELNUT	036	2025-06	1
C9840	CONTROL SERUM 88000-F17 HAZELNUT	036	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	3	9	0	3	6	6	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

639036651958277800
Hazelnut
CALIBRATOR 0.280 mL IVD
LOT 036
 2025-06 2°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	3	9	0	3	6	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

639036627500000001
Hazelnut
CONTROL 0.480 mL IVD
LOT 036
 2025-06 2°C
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents
DATE 27/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.