



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 033

2025-05

2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2222	Herpes simplex 1+2 IgM	033	2025-05	6x6
B2222	CALIBRATORE Herpes simplex 1+2 IgM	033	2025-05	1
C2222	CONTROL SERUM Herpes simplex 1+2 IgM	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	≥ 100
Control Serum (INDEX)	2,2	1,7-3,2
Negative Control (INDEX)	0,1	$< 0,7$

CALIBRATOR


1	5	6	0	3	3	6	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM


1	5	6	0	3	3	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

27/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.