

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG - CHORUS

Code	81040		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2271	TOXOPLASMA IgG	033	2025-05	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	033	2025-05	1
C2271	CONTROL SERUM TOXOPLASMA IgG	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (IU/mL)	79,5	31,3-100,0
Negative Control (IU/mL)	<5,0	<5

CALIBRATOR

0	8	2	0	3	3	6	1	1	7	2	9	3	5	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Oschi*

082033611729353800
Toxoplasma IgG
CALIBRATOR 0,425 mL IVD
LOT 033
2025-05 2°C-8°C
DIESSE Monteriggioni Italy 0123

CONTROL SERUM

0	8	2	0	3	3	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Oschi*

082033620380000001
Toxoplasma IgG
CONTROL 0,425 mL IVD
LOT 033
2025-05 2°C-8°C
DIESSE Monteriggioni Italy 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 26-05-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.