

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-M - CHORUS

Code 86034

Lot No. 033



2025-08



2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D7402	dsDNA-M	033	2025-08	6X6
B7402	CALIBRATORE dsDNA-M	033	2025-08	1
C7402	CONTROL SERUM dsDNA-M	033	2025-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	590	≥ 100
Control Serum (AU/mL)	68,5	55,9-94,1

CALIBRATOR

2	5	7	0	3	3	7	3	1	4	3	1	4	5	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM

2	5	7	0	3	3	7	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 28.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.