

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA IgA - CHORUS

Code	81033		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D7552	MYCOPLASMA IgA	033	2025-05	6X6
B7552	CALIBRATORE MYCOPLASMA IgA	033	2025-05	1
C7552	CONTROL SERUM MYCOPLASMA IgA	033	2025-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	550	≥ 100
Control Serum (AU/mL)	25,4	15,0-35,6
Negative Control (AU/mL)	<10,0	<10

CALIBRATOR

0	6	8	0	3	3	6	8	1	2	2	1	6	8	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *Open*



CONTROL SERUM

0	6	8	0	3	3	6	1	0	1	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *Open*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *28.05.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.