

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME INTRINSIC FACTOR - CHORUS

Code 86090

Lot No. 033

2025-05

2024-05

## REAGENTS

Label	Description	Lot.	Exp.	NO
D743I	INTRINSIC FACTOR	033	2025-05	6X6
B743I	CALIBRATORE INTRINSIC FACTOR	033	2025-05	1
C743I	CONTROL SERUM INTRINSIC FACTOR	033	2025-05	1

## CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	580	$\geq 100$
Control Serum (AU/mL)	76,6	38,3-100,0

## CALIBRATOR

5	1	2	0	3	3	6	3	1	9	2	9	4	5	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



## CONTROL SERUM

5	1	2	0	3	3	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 29.05.2024

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.