

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-T11 PLANE

Code	88000-T11		
Lot No.	035	2024-11	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D1121	88000-T11 PLANE	035	2024-11	2X6
B1120	CALIBRATORE 88000-T11 PLANE	035	2025-06	1
C1120	CONTROL SERUM 88000-T11 PLANE	035	2025-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	7	8	0	3	5	4	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

6	7	8	0	3	5	4	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]
DATE 29/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.