

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F9 RICE

Code	88000-F9		
Lot No.	036	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D6921	88000-F9 RICE	036	2025-05	2x6
B6920	CALIBRATORE 88000-F9 RICE	036	2025-06	1
C6920	CONTROL SERUM 88000-F9 RICE	036	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	4	6	0	3	6	6	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

646036651958277800
Rice
CALIBRATOR 0.280 mL IVD
LOT 036
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	4	6	0	3	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

646036627500000001
Rice
CONTROL 0.450 mL IVD
LOT 036
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 29/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.