

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-K82 LATEX

Code	88000-K82		
Lot No.	036	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D6941	88000-K82 LATEX	036	2025-05	2X6
B6940	CALIBRATORE 88000-K82 LATEX	036	2025-06	1
C6940	CONTROL SERUM 88000-K82 LATEX	036	2025-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	5	3	0	3	6	6	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

6	5	3	0	3	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 29/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.