



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-USTEKINUMAB

Code 86713  
Lot No. 001 2025-01 2024-05

REAGENTS

Label	Description	Lot	Exp.	NO
D4510	PROMONITOR ANTI-USTEKINUMAB	001	2025-01	6x6
B4510	CALIBRATORE ANTI-USTEKINUMAB	001	2025-07	1
C4510	CONTROL SERUM ANTI-USTEKINUMAB	001	2025-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	315	≥ 100
Control Serum (AU/mL)	16,2	13,9-21,1

CALIBRATOR

2 8 1 0 0 1 6 8 1 9 3 7 3 1 6 9 0 0

Label checked with barcode reader YES  Operator Signature *Abuwa*

281001681937316900  
Promonitor ANTI-USTEKINUMAB  
CALIBRATOR 0.850 mL IVD  
LOT 001  
2025-07 2°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2 8 1 0 0 1 6 2 6 3 9 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature *Abuwa*

281001626390000001  
Promonitor ANTI-USTEKINUMAB  
CONTROL 1.100 mL IVD  
LOT 001  
2025-07 2°C  
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Abuwa*

DATE 30.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.