Tel. +39 0577319554 FAX +39 0577319020

DIESSE DIAGNOSTICA SENESE SpA 53035 Monteriggioni (SI), Italy, Strada dei Laghi 39



Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

# QUALITY CONTROL CERTIFICATE

#### KIT NAME CHORUS PHOSPHOLIPID SCREEN IGM

Code 86108

Lot No. 033

2025-05

M 2024-05

#### REAGENTS

Label	Description	Lot.	Exp.	NO
D4020	Phospholipid Screen IgM	033	2025-05	6X6
B4020	CALIBRATORE Phospholipid Screen IgM	033	2025-05	1
C4020	CONTROL SERUM Phospholipid Screen IgM	033	2025-05	1

### CONTROLS

Found results   Expected result	
	_

Calibrator (mOD)	530	≥ 100			
Control (AU/ml)	21,2	14,3-35,6			

### CALIBRATOR

į,																		
ı	0	7	Е	_	7	7	_	1	1	-7	_	_	г	_	-	2		
١	9	/	0	U	ادا	3	Ь	1	1	/ /	4	٥١	5	2	5		l O	0

Label checked with barcode reader

YES 🔀

Operator Signature

ozu



# CONTROL SERUM

-							/				-						
9	7	5	0	3	3	6	1	3	2	4	0	0	0	0	0	0	1

Label checked with barcode reader

YES 🗶

Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 30.05-2024

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention