

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074/12

Lot No. 033



2025-05



2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	033	2025-05	2x6
B3591	CALIBRATORE a-TPO	033	2025-05	1
C3591	CONTROL SERUM a-TPO	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	≥ 100
Control Serum (IU/mL)	387,3	188,2-568,2

CALIBRATOR

3	0	1	0	3	3	6	6	2	4	4	1	3	0	6	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

3	0	1	0	3	3	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

30/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.