

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554

DIESSE

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA A IgG -CHORUS

Code 81190

Lot No. 033

2025-05

2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D7512	INFLUENZA A IgG	033	2025-05	6X6
B7512	CALIBRATORE INFLUENZA A IgG	033	2025-05	1
C7512	CONTROL SERUM INFLUENZA A IgG	033	2025-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (Index)	3,1	1,9-3,6
Negative Control (Index)	0,0	$\leq 0,7$

CALIBRATOR


0 7 0 0 3 3 6 0 0 0 0 0 3 6 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM

0 7 0 0 3 3 6 1 9 3 6 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *HB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

[Signature]

DATE 03/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.