



QUALITY CONTROL CERTIFICATE

KIT NAME Q FEVER PHASE II IgG - CHORUS

Code 81171

Lot No. 033 2025-05 2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2080	Q FEVER PHASE II IgG	033	2025-05	6x6
B2080	CALIBRATORE Q FEVER PHASE II IgG	033	2025-05	1
C2080	CONTROL SERUM Q FEVER PHASE II IgG	033	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	302	≥ 100
Control Serum (Index)	3,2	2,5-4,7

CALIBRATOR

9 4 4 0 3 3 6 0 0 0 0 0 2 8 0 0 0 0

Label checked with barcode reader YES Operator Signature

94403360000280000
Q Fever phase II IgG
CALIBRATOR 0.175 mL IVD
LOT 033
2025-05 2°C 8°C
DIESSSE Monteriggioni Italy CE
E2080

CONTROL SERUM

9 4 4 0 3 3 6 2 5 4 7 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature

94403362547000001
Q Fever phase II IgG
CONTROL 0.425 mL IVD
LOT 033
2025-05 2°C 8°C
DIESSSE Monteriggioni Italy CE
C2080

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 04-06-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.