



## EU DECLARATION OF CONFORMITY

MANUFACTURER:

DIESSE DIAGNOSTICA SENESE SPA  
STRADA DEI LAGHI 39  
53035 MONTERIGGIONI (SI),  
ITALY

SINGLE REGISTRATION NUMBER

IT-MF-000013311

AUTHORIZED REPRESENTATIVE:

NOT APPLICABLE

PRODUCT:  
CODE:

**CHORUS CALPROTECTIN 3RD GENERATION  
86603**

INTENDED PURPOSE:

CHORUS CALPROTECTIN 3RD GENERATION is an immunoassay kit for the automated quantitative detection of Calprotectin, performed in human stool using a disposable device applied on the CHORUS/ CHORUS TRIO instruments. High concentration of Calprotectin is observed in the stool of patients with chronic inflammatory diseases (IBD). The kit, intended for the determination of Calprotectin concentration is used as an aid to diagnosis and to allow monitor of related therapeutic measures. It is to be used only by professional laboratory personnel.

BASIC UDI-DI

803389132CHORUSCALP3009M

UDI-DI

08033891328436

RISK CLASS:

CLASS B

CLASSIFICATION RULE:

RULE 6

CONFORMITY ASSESSMENT ROUTE:

ANNEX IX

WE HEREWITH DECLARE THAT THE ABOVE MENTIONED PRODUCT IS IN CONFORMITY WITH **REGULATION (EU) 2017/746** OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON IN VITRO DIAGNOSTIC MEDICAL DEVICE. ALL SUPPORTING DOCUMENTATION IS RETAINED UNDER THE PREMISES OF THE MANUFACTURER.

REFERENCE TO ANY CS APPLIED:

NOT APPLICABLE

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH  
RIDLERSTAßE 65 – 80339 MUNICH  
GERMANY  
No. 0123

(EU) CERTIFICATE:

V12 056726 0006 Rev. 02

REVISION:

1

PLACE, DATE OF ISSUE:

MONTERIGGIONI, 05 JUNE 2024

EXPIRY DATE:

2027-06-26

THE PRESENT EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



MARIA CLAUDIA ALCARO  
PERSON RESPONSIBLE FOR THE REGULATORY  
COMPLIANCE

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 2024-06-05



MAGDALENA STOCZKO  
REGULATORY SUPERVISOR