



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554

QUALITY CONTROL CERTIFICATE

KIT NAME 88000- I1- HONEY BEE

Code	88000-I1		
Lot No.	036	2025-05	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D6931	88000- I1 HONEY BEE	036	2025-05	2x6
B6930	CALIBRATORE 88000- I1 HONEY BEE	036	2025-06	1
C6930	CONTROL SERUM 88000- I1 HONEY BEE	036	2025-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,7	6,6-15,0

CALIBRATOR

6	4	5	0	3	6	6	5	1	9	5	8	2	7	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

6	4	5	0	3	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 05/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.