

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA URINARY ANTIGEN - CHORUS

Code	81301
Lot No.	033
	2025-02
	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9530	LEGIONELLA URINARY ANTIGEN	033	2025-02	2X6
B9530	CALIBRATOR LEGIONELLA URINARY ANTIGEN	033	2025-02	1
C9530	POSITIVE CONTROL LEGIONELLA URINARY ANTIGEN	033	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	220	≥ 100
Positive Control (Index)	3,2	2,3-4,3

CALIBRATOR

2	7	0	0	3	3	5	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

270033500000300000  
Legionella Urinary  
Antigen  
CALIBRATOR 0.350 mL IVD  
LOT 033  
2025-02  
DIESSSE Monteriggioni Italy  
CE  
B9530

CONTROL SERUM

2	7	0	0	3	3	5	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

270033523430000001  
Legionella Urinary  
Antigen  
CONTROL+ 0.850 mL IVD  
LOT 033  
2025-02  
DIESSSE Monteriggioni Italy  
CE  
C9530

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 07.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.