



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 035 2025-05 2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	035	2025-05	6X6
B3411	CALIBRATORE Gliadin-G	035	2025-05	1
C3411	CONTROL SERUM Gliadin-G	035	2025-05	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	569	≥ 100
CONTROL SERUM (AU/ml)	40,0	25,8-62,5

CALIBRATOR

5 0 2 0 3 5 6 2 2 3 3 5 5 0 4 2 0 0

Label checked with barcode reader YES Operator Signature *Goleen*



CONTROL SERUM

5 0 2 0 3 5 6 1 4 2 6 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *Goleen*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 06.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.