

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME CHIKUNGUNYA IgM -CHORUS

Code 81134

Lot No. 024



2025-04



2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D0211	CHIKUNGUNYA IgM	024	2025-04	6x6
B0211	CALIBRATORE CHIKUNGUNYA IgM	024	2025-05	1
C0211	CONTROL CHIKUNGUNYA IgM	024	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	457	≥ 100
Control (INDEX)	2,1	1,4-2,6

CALIBRATOR

9	1	0	0	2	4	6	0	0	0	0	0	5	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

9	1	0	0	2	4	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 12.06.24

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.