

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

Code	81020
Lot No.	033
	2025-08
	2024-05

REAGENTS

Label	Description	Lot	Exp.	NO
D2212	Herpes simplex 1+2 IgG	033	2025-08	6X6
B2213	CALIBRATORE Herpes simplex 1+2 IgG	033	2025-08	1
C2212	CONTROL SERUM Herpes simplex 1+2 IgG	033	2025-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	429	≥ 100
Control Serum (INDEX)	3,2	2,8-5,2
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	0	4	0	3	3	7	0	0	0	0	0	2	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

0	0	4	0	3	3	7	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *12.06.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.