

DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME **88000-F25 TOMATO**

|         |           |         |         |
|---------|-----------|---------|---------|
| Code    | 88000-F25 |         |         |
| Lot No. | 042       | 2025-06 | 2024-06 |

REAGENTS

| Label | Description                    | Lot | Exp.    | NO  |
|-------|--------------------------------|-----|---------|-----|
| D9851 | 88000-F25 TOMATO               | 042 | 2025-06 | 2x6 |
| B9850 | CALIBRATORE 88000-F25 TOMATO   | 042 | 2025-08 | 1   |
| C9850 | CONTROL SERUM 88000-F25 TOMATO | 042 | 2025-08 | 1   |

CONTROLS

|                      | Found results | Expected results |
|----------------------|---------------|------------------|
| Calibrator (mOD)     | 220           | ≥ 100            |
| Control Serum (KU/L) | 11,4          | 5,5-15,6         |

CALIBRATOR

6 1 3 0 4 2 6 5 1 9 4 9 3 0 8 2 0 0

Label checked with barcode reader YES  Operator Signature *FB*

613042651949308200  
Tomato  
CALIBRATOR 0.250 mL IVD  
LOT 042  
 2025-08 2°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6 1 3 0 4 2 6 2 3 4 4 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature *FB*

613042623440000001  
Tomato  
CONTROL 0.450 mL IVD  
LOT 042  
 2025-08 2°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *12/06/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.