

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code 86060

Lot No. 033



2025-05



2024-06

REAGENTS

Label	Description	Lot	Exp.	NO
D3452	tTg-G	033	2025-05	6X6
B3452	CALIBRATORE tTg-G	033	2025-05	1
C3452	CONTROL SERUM tTg-G	033	2025-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	344	≥ 100
Control Serum (AU/mL)	44,7	31,6-56,8

CALIBRATOR

3	4	0	0	3	3	6	1	1	8	3	7	3	7	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

3	4	0	0	3	3	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *FB*

DATE 12/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.