

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	033	2025-05	6X6
B3281	CALIBRATORE dsDNA-G	033	2025-05	1
C3281	CONTROL SERUM dsDNA-G	033	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	390	≥ 100
Control Serum (IU/mL)	60,2	39,1-82,1

CALIBRATOR

2	5	1	0	3	3	6	6	2	6	4	0	3	9	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

251033662640394400
dsDNA-G
CALIBRATOR 0.176 mL [IVD]
LOT 033
2025-05 2°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	1	0	3	3	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

251033619350000001
dsDNA-G
CONTROL 0.425 mL [IVD]
LOT 033
2025-05 2°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 14.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.