



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554

QUALITY CONTROL CERTIFICATE

KIT NAME ASCA-G - CHORUS

Code	86064		
Lot No.	043	2025-06	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3491	ASCA-G	043	2025-06	6X6
B3491	CALIBRATORE ASCA-G	043	2025-06	1
C3491	CONTROL SERUM ASCA-G	043	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	427	≥ 100
Control Serum (AU/mL)	84,2	50,0-123,8

CALIBRATOR

5	0	8	0	4	3	6	4	2	0	3	1	3	6	7	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

508043642031367300
ASCA-G
CALIBRATOR 0.176 mL IVD
LOT 043
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

5	0	8	0	4	3	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

508043620360000001
ASCA-G
CONTROL 0.425 mL IVD
LOT 043
2025-06 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 27/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.