



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** CYTOMEGALOVIRUS IgG -CHORUS

Code	81010		
Lot No.	016	2025-03	2024-03

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	016	2025-03	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	016	2025-03	1
C2161	CONTROL SERUM Cytomegalovirus IgG	016	2025-03	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	310	≥ 100
Control Serum (IU/mL)	5,8	3,7-10,0
Negative Control (IU/mL)	<0,5	<0,5

**CALIBRATOR**

0	0	1	0	1	6	6	5	2	8	4	1	3	2	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

001016652841326200  
Cytomegalovirus IgG  
CALIBRATOR 0.425 mL IVD  
LOT 016  
 2025-03 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123 B2162

**CONTROL SERUM**

0	0	1	0	1	6	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

001016622410000001  
Cytomegalovirus IgG  
CONTROL 0.425 mL IVD  
LOT 016  
 2025-03 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123 C2161

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]  
DATE 19/04/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.