

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code 81041

Lot No. 018 2025-03 2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	018	2025-03	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	018	2025-03	1
C2281	CONTROL SERUM TOXOPLASMA IgM	018	2025-03	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	413	≥ 100
Control Serum (INDEX)	2,1	1,6-3,0
Negative Control (INDEX)	0,1	<0,7

CALIBRATOR

1 3 1 0 1 8 6 0 0 0 0 0 4 4 0 0 0 0


Label checked with barcode reader YES Operator Signature Ben



CONTROL SERUM

1 3 1 0 1 8 6 1 6 3 0 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature Ben



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Ben

DATE 18-06-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.