

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554

DIESSE

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgM CHORUS

Code 81166

Lot No. 042



2025-06



2024-06

REAGENTS

Label	Description	Lot	Exp.	NO
D4120	Brucella IgM	042	2025-06	6X6
B4120	CALIBRATORE Brucella IgM	042	2025-06	1
C4120	Control Serum Brucella IgM	042	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	549	≥ 100
Control (Index)	2,0	1,4-2,5

CALIBRATOR

1	4	7	0	4	2	6	0	0	0	0	0	5	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

MB



CONTROL SERUM

1	4	7	0	4	2	6	1	4	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

MB



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 20/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.