

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



### QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code 81021

Lot No. 035

2025-05

2024-05

### REAGENTS

| Label | Description                          | Lot. | Exp.    | NO  |
|-------|--------------------------------------|------|---------|-----|
| D2222 | Herpes simplex 1+2 IgM               | 035  | 2025-05 | 6x6 |
| B2222 | CALIBRATORE Herpes simplex 1+2 IgM   | 035  | 2025-05 | 1   |
| C2222 | CONTROL SERUM Herpes simplex 1+2 IgM | 035  | 2025-05 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |     |            |
|--------------------------|-----|------------|
| Calibrator (mOD)         | 420 | $\geq 100$ |
| Control Serum (INDEX)    | 2,1 | 1,4-2,6    |
| Negative Control (INDEX) | 0,1 | $< 0,7$    |

### CALIBRATOR


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 5 | 6 | 0 | 3 | 5 | 6 | 0 | 0 | 0 | 0 | 0 | 5 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *[Signature]*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 5 | 6 | 0 | 3 | 5 | 6 | 1 | 4 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

*[Signature]*

DATE 19-06-2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.