

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



## QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 043



2025-06



2024-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3142	ANA-Screen	043	2025-06	6X6
B3142	CALIBRATORE ANA-Screen	043	2025-06	1
C3142	CONTROL SERUM ANA-Screen	043	2025-06	1

### CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	355	$\geq 100$
Control Serum (Ndx)	2,4	1,8-3,3

### CALIBRATOR

1	1	5	0	4	3	6	0	0	0	0	0	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



### CONTROL SERUM

1	1	5	0	4	3	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *20/06/2024*

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.