



## QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code 81059

Lot No. 042



2025-06



2024-06

### REAGENTS

Label	Description	Lot	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	042	2025-06	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	042	2025-06	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	042	2025-06	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	410	$\geq 100$
Control Serum (Index)	2,2	1,8-3,3
Negative Control (Index)	0,0	$< 0,7$


### CALIBRATOR

0	2	9	0	4	2	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 





### CONTROL SERUM

0	2	9	0	4	2	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 20.06.2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.