

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-W9 PLAINTAIN

Code	88000-W9		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D5711	88000-W9 PLAINTAIN	042	2025-06	2X6
B5710	CALIBRATORE 88000-W9 PLAINTAIN	042	2025-08	1
C5710	CONTROL SERUM 88000-W9 PLAINTAIN	042	2025-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	220	≥ 100
Control Serum (KU/L)	11,4	5,5-15,6

CALIBRATOR

6	5	4	0	4	2	6	5	1	9	4	9	3	0	8	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

654042651949308200
Plaintain
CALIBRATOR 0.250 mL IVD
LOT 042
2025-08 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	5	4	0	4	2	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

654042623440000001
Plaintain
CONTROL 0.450 mL IVD
LOT 042
2025-08 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 21/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.