

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-W5 WORMWOOD

Code	88000-W5		
Lot No.	042	2025-06	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D4991	88000-W5 WORMWOOD	042	2025-06	2X6
B4990	CALIBRATORE 88000-W5 WORMWOOD	042	2025-08	1
C4990	CONTROL SERUM 88000-W5 WORMWOOD	042	2025-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	220	≥ 100
Control Serum (KU/L)	11,4	5,5-15,6

626042651949308200
Wormwood

CALIBRATOR 0.280 mL IVD

LOT 042

2025-08 2°C B4990

DIESE Monteriggioni Italy

CALIBRATOR

6	2	6	0	4	2	6	5	1	9	4	9	3	0	8	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

CONTROL SERUM

6	2	6	0	4	2	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

626042623440000001
Wormwood

CONTROL 0.480 mL IVD

LOT 042

2025-08 2°C C4990

DIESE Monteriggioni Italy

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 24/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.