

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code	86098		
Lot No.	027	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D8651	Deamidated Gliadin-A	027	2025-04	6X6
B8651	CALIBRATORE Deamidated Gliadin-A	027	2025-04	1
C8651	CONTROL SERUM Deamidated Gliadin-A	027	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	400	≥ 100
Control Serum (AU/mL)	140,4	67,9-284,6

CALIBRATOR

2	5	4	0	2	7	6	2	1	9	3	3	2	9	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

254027621933294600
DEAMIDATED GLIADIN-A
CALIBRATOR 0.175 mL IVD
LOT 027
2025-04
2°C-8°C
DIESSE Monteriggioni Italy
B8651
CE 0123

CONTROL SERUM

2	5	4	0	2	7	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

254027624450000001
DEAMIDATED GLIADIN-A
CONTROL 0.425 mL IVD
LOT 027
2025-04
2°C-8°C
DIESSE Monteriggioni Italy
C8651
CE 0123

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Beura*

DATE 21.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.