

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



## QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 027



2025-07



2024-04

### REAGENTS

Label	Description	Lot	Exp.	NO
D2431	MEASLES IgG	027	2025-07	6X6
B2432	CALIBRATORE MEASLES IgG	027	2025-07	1
C2431	CONTROL SERUM MEASLES IgG	027	2025-07	1

### CONTROLS

	Found results	Expected results
Calibrator (mOD)	426	$\geq 100$
Control Serum (Index)	3,2	2,2-4,1
Negative Control (Index)	0,0	$< 0,7$

### CALIBRATOR

0	0	9	0	2	7	7	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature ..... *FB*




### CONTROL SERUM

0	0	9	0	2	7	7	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature ..... *FB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 21/06/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.