

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada del Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG - CHORUS

Code	81040		
Lot No.	034	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2271	TOXOPLASMA IgG	034	2025-05	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	034	2025-05	1
C2271	CONTROL SERUM TOXOPLASMA IgG	034	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control Serum (IU/mL)	40,4	30,8-78,6
Negative Control (IU/mL)	<5,0	<5

CALIBRATOR

0	8	2	0	3	4	6	2	1	7	3	0	3	6	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

082034621730364400  
Toxoplasma IgG  
CALIBRATOR 0.425 mL IVD  
LOT 034  
2025-05 2°C 8°C B2272  
DIESSE Monteriggioni Italy CE 0123

CONTROL SERUM

0	8	2	0	3	4	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

082034620380000001  
Toxoplasma IgG  
CONTROL+ 0.425 mL IVD  
LOT 034  
2025-05 2°C 8°C C2271  
DIESSE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 25/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.