

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME ADENOVIRUS IgG -CHORUS

Code 81196

Lot No. 025

2025-04

2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D7601	Adenovirus IgG	025	2025-04	6X6
B7601	CALIBRATORE Adenovirus IgG	025	2025-04	1
C7601	CONTROL SERUM Adenovirus IgG	025	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	≥ 100
Control Serum (Index)	1,7	1,3-2,4
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	7	6	0	2	5	6	0	0	0	0	0	5	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*




CONTROL SERUM

0	7	6	0	2	5	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 25-06-2026

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.