

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554

DIESSE

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270

Lot No. 027

2025-04

2024-04

REAGENTS

Label	Description	Lot	Exp.	NO
D9660	Tetanus IgG	027	2025-04	EX6
B9660	CALIBRATORE Tetanus IgG	027	2025-04	1
C9660	CONTROL SERUM Tetanus IgG	027	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control (mIU/mL)	649,3	500,0-1157,9


CALIBRATOR

9 0 2 0 2 7 6 3 1 9 3 2 3 8 5 1 0 0

Label checked with barcode reader

YES ☒

Operator Signature *Blue*




CONTROL SERUM

9 0 2 0 2 7 6 1 9 3 4 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *Blue*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE 25.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.