



QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS LISTERIA MONOCYTOGENES CONTROL SERUM

Code	85875		
LOT.NO	083	2025-06	2024-06

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF90875	B2980	LISTERIA MONOCYTOGENES CONTROL SERUM	083	2025-06	1

CONTROLS

TESTED WITH:	FOUND RESULTS	EXPECTED RESULTS
CHORUS LISTERIA MONOCYTOGENES LOT 017	1/6	1/16± 1 Dil.

Evaluation of the method according to the package insert in the attached document  
This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 26.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIALI

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.