

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM CAPTURE - CHORUS

Code	81042
Lot No.	042
	2025-06
	2024-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3900	TOXOPLASMA IgM CAPTURE	042	2025-06	6X6
B3900	CALIBRATORE TOXOPLASMA IgM CAPTURE	042	2025-06	1
C3900	CONTROL SERUM TOXOPLASMA IgM CAPTURE	042	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	378	≥ 100
Control Serum (Index)	3,0	1,7-3,2

CALIBRATOR

9	4	9	0	4	2	6	0	0	0	0	0	4	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	4	9	0	4	2	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 26/06/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.