

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554



### QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code 86601

Lot No. 042



2025-06



2024-06

### REAGENTS

| Label | Description              | Lot. | Exp.    | NO  |
|-------|--------------------------|------|---------|-----|
| D9970 | CALPROTECTIN             | 042  | 2025-06 | 2X6 |
| B9970 | CALIBRATORE CALPROTECTIN | 042  | 2025-10 | 1   |
| C9970 | CONTROL CALPROTECTIN     | 042  | 2025-10 | 1   |
| G9502 | SAMPLE DILUENT           | 042  | 2026-02 | 1   |

### CONTROLS

| Found results | Expected results |
|---------------|------------------|
|---------------|------------------|

|                       |       |              |
|-----------------------|-------|--------------|
| Calibrator (mOD)      | 300   | $\geq 100$   |
| Control Serum (mg/Kg) | 647,6 | 424,1-1000,0 |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 2 | 9 | 0 | 4 | 2 | 6 | 4 | 2 | 5 | 5 | 4 | 2 | 1 | 6 | 2 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 2 | 9 | 0 | 4 | 2 | 6 | 3 | 4 | 6 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 28/06/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.